



Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

OCT 29 2002

Mr. David Olmstead  
Vice President, RA/QA  
Futura Medical Technology  
2025 W. 11<sup>th</sup> Street  
Upland, California 91786

Re: K023323  
Trade/Device Name: Futura® Safety Syringe  
Regulation Number: 880.5860  
Regulation Name: Piston Syringe  
Regulatory Class: II  
Product Code: MEG  
Dated: October 3, 2002  
Received: October 4, 2002

Dear Mr. Olmstead:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies.

You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 21 CFR Part 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4618. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,



Timothy A. Ulatowski  
Director

Division of Anesthesiology, General Hospital,  
Infection Control and Dental Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

Section 2

**Indications for Use Statement**

510(k) Number            Not applicable

Device Name:            Futura® Safety Syringe

Indications For Use      The device is intended to be used for subcutaneous and intramuscular use

The function of the Futura® Safety Syringe is designed to provide safe and reliable method of injecting medication and fluids into patients and also helps protect the healthcare worker from potential needle sticks.

The Futura® Safety Syringe functions as a conventional hypodermic syringe except for its ability to retract the contaminated needle safely inside the syringe immediately after the completion of the patient injection. Complete delivery of the syringe contents activates the retraction mechanism.

(PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF NEEDED)

\_\_\_\_\_ Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use ✓ or Over-The-Counter Use \_\_\_\_\_  
(Per 21 CFR 801)

*Josh Holland for Pat Criventi*  
(Division Sign-Off)  
Division of Anesthesiology, General Hospital,  
Infection Control, Dental Devices

510(k) Number: K023323